



## Alanine Aminotransferase-SL (ALT-SL)

FOR THE QUANTITATIVE MEASUREMENT OF ALT

METHOD: ENZYMATIC (MODIFIED IFCC); KINETIC

ALT measurements are used as an aid to monitor and diagnose liver diseases.

The SEKISUI ALT-SL assay incorporates optimized substrate concentrations with a tris buffer. It is intended for the measurement of ALT in serum.

### Features

- Two part, stable liquid
- Wide variety of chemistry analyzer applications are available

### Benefits

- Easy to use, no additional reagent preparation required
- Flexible testing, well suited for use with fully automated procedures

### Performance Characteristics

#### Precision

- Within-Run:  $\leq 2.4\%$
- Total Precision:  $\leq 3.6\%$

#### Accuracy<sup>(a)</sup>

- Slope: 1.06
- Intercept: -2.8 U/L
- Correlation Coefficient: 0.9945

#### Linearity

- 10 - 600 U/L

(a) The performance of this method (y) was compared with the performance of a similar ALT method (x) on a Roche/Hitachi® analyzer.

#### No Significant Interferences Up to Levels Indicated

- Bilirubin: 40 mg/dL (684  $\mu\text{mol/L}$ )
- Intralipid: 400 mg/dL (1200 mg/dL (13.6 mmol/L) Simulated Triglycerides)

#### Reference Range<sup>(1)</sup>

- Newborn: 7- 40 U/L (37°C)
- Adult: 10 - 35 U/L (37°C)

(1) Burtis, C.A., Ashwood, E.R. (Ed.), Tietz Textbook of Clinical Chemistry, W.B. Saunders Co., Toronto, p. 790-791, 797 (1994).

# Alanine Aminotransferase-SL (ALT-SL)


## Ordering Information

	Configuration	Catalog Number
ALT-SL	R1 3 x 100mL R2 1 x 75mL	318-30
DC-CAL CALIBRATOR	5 x 3mL	SE-035
DC-TROL LEVELS 1 AND 2	Level 1: 5 x 5mL Level 2: 5 x 5mL	SM-057

To learn more about our clinical chemistry reagent products scan the QR code or go to [Clinical Chemistry Reagents – SEKISUI Diagnostics](#).



Sam



Skyler



Suzy

Securing accurate and reliable Clinical Chemistry results!



### THE AMERICAS

SEKISUI Diagnostics, LLC  
One Wall Street  
Burlington, MA 01803  
Phone: 800 332 1042

Email: [questions@sekisui-dx.com](mailto:questions@sekisui-dx.com)

### INTERNATIONAL

Sekisui Diagnostics (UK) Limited  
Liphook Way, Allington  
Maidstone, Kent, ME16 0LQ, UK  
Phone: +44 1622 607800

Email: [info@sekisui-dx.com](mailto:info@sekisui-dx.com)

**SEKISUI**  
DIAGNOSTICS

Because every result matters™

[sekisuidiagnostics.com](http://sekisuidiagnostics.com)