



Calcium

FOR THE QUANTITATIVE MEASUREMENT OF CALCIUM

METHOD: ARSENAZO III; ENDPOINT

Calcium measurements are used as an aid to monitor and diagnose bone and kidney disorders as well as neurological disease and pancreatitis.

The SEKISUI Calcium assay uses an Arsenazo III procedure which is easy to use and applicable to routine laboratory instrumentation. It is intended for the measurement of calcium in serum.

Features	
 No significant interference displayed from samples with elevated levels of hemolysis, lipemia or icterus One part, stable liquid 	Wide variety of chemistry analyzer applications available
Benefits	
 Reduces the need for sample dilutions Easy to use, no additional preparation required 	 Confidence in results Flexible testing, well suited for use with fully automated procedures
Performance Characteristics	
Precision Within-Run: ≤0.8% Total Precision: ≤1.0% Accuracy ^(a) Slope: 0.875 Intercept: 0.91 mg/dL (0.23 mmol/L) Correlation Coefficient: 0.9889 Linearity 0.04 -15.00 mg/dL (0.01- 3.75 mmol/L)	No Significant Interferences Up to Levels Indicated Hemoglobin: 1000 mg/dL (155 µmol/L) Bilirubin: 40 mg/dL (684 µmol/L) Intralipid: 1000 mg/dL (3000 mg/dL (33.9 mmol/L) Simulated Triglycerides) Reference Range(1) SERUM Premature Infants: 6.0 - 10.0 mg/dL (1.5 -2.5 mmol/L) Full-term infants: 7.3 - 12.0 mg/dL (1.8- 3.0 mmol/L) 1 to 2 years: 10.0 - 12.0 mg/dL (2.5- 3.0 mmol/L) Adults: 8.0 - 10.5 mg/dL (2.0 - 2.6 mmol/L)

⁽a) The performance of this method (y) was compared with the performance of another calcium method (x).

⁽¹⁾ Kaplan, Lawrence A. and Pesce, Amadeo J., (Eds.), Clinical Chemistry: Theory, Analysis, Correlation, Mosby-Year Book Inc., St. Louis (1996). *Calibrator for Manual Assays Only



Calcium

Ordering Information		
	Configuration	Catalog Number
CALCIUM	R1 4 x 125mL (Reagent) 1 x15mL (Calibrator)	140-20
DC-CAL CALIBRATOR	5 x 3mL	SE-035
DC-TROL LEVELS 1 & 2	Level 1 5 x 5mL Level 2 5 x 5mL	SM-057

To learn more about our clinical chemistry reagent products scan the QR code or go to Clinical Chemistry Reagents - SEKISUI Diagnostics.





Securing accurate and reliable Clinical Chemistry results!



THE AMERICAS

SEKISUI Diagnostics,LLC One Wall Street Burlington, MA 01803 Phone: 800 332 1042 Email: questions@sekisui-dx.com

INTERNATIONAL

Sekisui Diagnostics (UK) Limited Liphook Way, Allington Maidstone, Kent, ME16 0LQ, UK Phone: +44 1622 607800 Email: info@sekisui-dx.com

