



Unsaturated Iron Binding Capacity (UIBC-SL)

FOR THE QUANTITATIVE MEASUREMENT OF UNSATURATED IRON BINDING CAPACITY (UIBC)

METHOD: FERENE®; ENDPOINT

UIBC measurements are used in combination with serum iron as an aid to diagnose and monitor various iron disorders.

The SEKISUI UIBC-SL assay uses a superior iron chelating agent forming a Ferene® complex with ferrous iron. It is intended for the measurement of unsaturated iron binding capacity in serum.

Features

- Two part stable liquid ready to use reagent
- Applicable to multiple chemistry platforms
- Wide linear range compared to the reference range

Benefits

- Easy to use, no additional reagent preparation required
- Flexible testing, well suited for use with fully automated procedures
- Reduces the need for diluting and repeat testing

Performance Characteristics

Precision

- Within-Run: $\leq 5.0\%$
- Total Precision: $\leq 5.7\%$

Accuracy^(a)

- Slope: 1.065
- Intercept: 2.0 mg/dL (0.4 $\mu\text{mol/L}$)
- Correlation Coefficient: 0.9937

Linearity

- 17 - 600 $\mu\text{g/dL}$ (3.0 - 107.4 $\mu\text{mol/L}$)

No Significant Interferences Up to Levels Indicated

- Bilirubin: 40 mg/dL (684 $\mu\text{mol/L}$)
- Intralipid: 1000 mg/dL (3000 mg/dL (33.9 mmol/L) Simulated Triglycerides)
- Hemoglobin: 400 mg/dL (62 $\mu\text{mol/L}$)

Reference Range

- 126 - 382 $\mu\text{g/dL}$ (22.6 - 68.4 $\mu\text{mol/L}$)


(a) The performance of this method (y) was compared with the performance of a similar UIBC method (x) on a Roche/Hitachi® analyzer.

Unsaturated Iron Binding Capacity (UIBC-SL)

Ordering Information

	Configuration	Catalog Number
UIBC-SL	R1 1 x 100mL R2 1 x 25mL	153-10
UIBC-SL	R1 3 x 100mL R2 1 x 75mL	153-30
UIBC-SL	R1 1 x 500mL R2 1 x 300mL	153-50
UIBC-SL	R1 1 x 1000mL R2 1 x 240mL	153-90
DC-UIBC-CAL CALIBRATION MATERIAL	6 x 10mL	SE-090
DC-LINEATE LINEARITY MATERIAL	LEVEL 1 2 X 5mL LEVEL 2 2 X 5mL LEVEL 3 2 X 5mL	SE-091

To learn more about our clinical chemistry reagent products scan the QR code or go to [Clinical Chemistry Reagents – SEKISUI Diagnostics](#).



Securing accurate and reliable Clinical Chemistry results!



THE AMERICAS

SEKISUI Diagnostics, LLC
One Wall Street
Burlington, MA 01803
Phone: 800 332 1042

Email: questions@sekisui-dx.com

INTERNATIONAL

Sekisui Diagnostics (UK) Limited
Liphook Way, Allington
Maidstone, Kent, ME16 0LQ, UK
Phone: +44 1622 607800

Email: info@sekisui-dx.com

SEKISUI
DIAGNOSTICS

Because every result matters™

sekisuidiagnostics.com