

ARK™ Gabapentin Control

This ARK Diagnostics, Inc. package insert for the ARK Gabapentin Control must be read carefully prior to use. Package insert instructions must be followed accordingly. Reliability of the assay results cannot be guaranteed if there are any deviations from the instructions in this package insert.











CUSTOMER SERVICE

 ARK Diagnostics, Inc.
48089 Fremont Blvd
Fremont, CA 94538 USA
Tel: 1-877-869-2320
Fax: 1-510-270-6298
customersupport@ark-tdm.com
www.ark-tdm.com



Emergo Europe
Prinsessegracht 20
2514 AP The Hague
The Netherlands

KEY TO SYMBOLS USED

	Batch code	 YYYY-MM-DD	Use by/Expiration date
	Catalog Number		Manufacturer
	Authorized Representative		CE Mark
	In Vitro Diagnostic Medical Device		Temperature limitation
	Consult Instructions for Use		Quality Control
Rx Only	For Prescription Use Only		

1 NAME

ARK™ Gabapentin Control

2 INTENDED USE

ARK Gabapentin Control is intended for use in quality control of the ARK Gabapentin Assay.

3 CONTENT

ARK Gabapentin Control is comprised of a synthetic protein matrix with the following concentrations of gabapentin:

REF	Product Description	Quantity/Volume
5025-0003-00	ARK Gabapentin Control* Gabapentin, buffer, bovine serum albumin, and sodium azide (nominal level)	Dropper vials
	LOW (2.5 µg/mL)	1 X 4 mL
	MID (8.0 µg/mL)	1 X 4 mL
	HIGH (25.0 µg/mL)	1 X 4 mL

* To convert results from µg/mL gabapentin to µmol/L gabapentin, multiply µg/mL by 5.84. Gabapentin levels become 14.6, 46.7, and 146.0 µmol/L for trilevel controls respectively.

Each laboratory should establish its own ranges for each new lot of controls.

4 WARNINGS AND PRECAUTIONS

- For **In Vitro Diagnostic** Use.
- Do not mix controls from different lot numbers.
- Use each lot as a set.
- Controls contain ≤0.09% sodium azide.

5 INSTRUCTIONS FOR USE

- For a complete summary and explanation of the gabapentin assay, refer to the package insert for the ARK Gabapentin Assay.
- Controls are ready to use. Mix each level by gentle inversion before dispensing.
- Squeeze sufficient volume (~40µL/drop) into individual sample cups for each level. Consult instrument-specific sample volume requirements. Return caps to their original containers and keep tight.
- Store at 2-8°C. Use prior to the expiration date. Once opened vials may be stored at 2-8°C for 12 months within the expiration date.

6 LIMITATIONS OF PROCEDURE

Accurate and reproducible results are dependent upon properly functioning instruments, reagents, calibrators, controls, storage of product as directed, and good laboratory technique.

7 TRADEMARKS

ARK™ is a trademark of ARK Diagnostics, Inc.

Other brand or product names are trademarks of their respective holders.